

Sunrise Instruments Calibration Report

Address: 2nd Floor Building B2 19#Keyuan Xishi Road Nanning Guangxi Province China 530007
Phone: +86 771-389-9499 Fax: +86 771-389-9497
Email: sri@srisensor.com



Calibration No.	SRI-OR-E005313	Cal Date	2026/1/22
Model No.	M2213A9	Serial No.	40181
Technician	Huang zeng ying	Temp(C)/Hum.(%)	24/67
Customer		Excitation(V)	5.0000
Description	TORQUE D60mm 100Nm	Cable Length	45mm
Comments	SHCS M5:8N.m SHCS M2.5:1N.m		

Voltage Calibration

Bridge	Capacity	Zero Offset	Nonlinearity	Hysteresis	Output @ Capacity	Sensitivity
	Nm	mV/V	%FS	%FS	mV/V	mV/V/EU
MZ	-100	-0.0445	0.33	0.25	-1.5980	1.5980E-02

产品型号	出厂SN	零点mV	灵敏度mV/Nm	折算灵敏度mV/V/Nm	发货日期	测试问题记录	备注
R75	40181	-0.22	0.0799	0.0160	2026/1/28		

Wire Color Codes

FUNCTION

+EXC
+SIG
-EXC
-SIG
Sensor ID
ID Return

Reference Load Cell

Manufacturer	Model Type	Model No.	Serial No.	Cal Due Date
SUNRISE	Golden Standard	SRI_2000N-1	30729	2026.05.11
Traceable to the National Institute of Metrology China (NIMC)				

This calibration follows the calibration document SRI-LAB-BZ-001:2024 which defined in the Sunrise Instruments Lab

The load cell calibration system consists of a 300KN load frame CMT5305 made by SANS Testing Machine Co., Golden reference load cell by Interface. Data acquisition system by National Instruments.

Calibrated by Huang zeng ying Approved by Andy

Sunrise Instruments Calibration Report

Address: 2nd Floor Building B2 19#Keyuan Xishi Road Nanning Guangxi Province China 530007
Phone: +86 771-389-9499 Fax: +86 771-389-9497
Email: sri@srisensor.com



Calibration No.	SRi-OR-E005313	Cal Date	2026/1/22
Model No.	M2213A9	Serial No.	40181
Technician	Huang zeng ying	Temp(C)/Hum.(%)	24/67
Customer		Excitation(V)	5.0000
Description	TORQUE D60mm 100Nm	Cable Length	45mm
Comments	SHCS M5:8N.m SHCS M2.5:1N.m		

CROSSTALK

<u>Bridge</u>	<u>Load</u>	<u>MZ</u>
	N/Nm	% FS
FX	-250	0.23
FY	250	0.21
FZ	-250	0.20
MX	50	0.48
MY	-50	0.39
MZ	-100	0.00

Bridge to Transducer body resistance greater than 5G Ohm when measured with 50V excitation

Page 2 of 2

_____blank below_____